

Lot Number: **B1925**

Reference Material Data

Organism: *Candida albicans*
 Designation: NCPF 3179 (ATCC 10231)
 Reference: 56013
 BTF Internal Reference: B-CA3179-550-20
 Date of Manufacture: 12 January 2012
 Expiry Date: 11 January 2014
 Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-019)

cfu Count per BioBall®

Mean cfu	SD
540.3	20.8

Quantification Data

cfu Count per 100µl aliquot*

Mean cfu	SD
49.1	5.5

95% Prediction Interval

Lower Limit	Upper Limit
37.9	60.3

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
 QC Media: Nutrient agar
 Plate Incubation Time (hrs): 40 - 48
 Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Officer:



.....
Linh Le

2 February 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer
 Accreditation Number: 14993
 Site Number: 15276

Strain sourced from NCPF (National Collection of Pathogenic Fungi)

ATCC is a trademark of American Type Culture Collection

*BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

* QC data using BioBall Re-Hydration Fluid (ref 56021)

Water Industry Enquires:

BTF Pty Ltd.
 Sydney Australia
 Tel. 61 (0) 2 8877 9150
 Fax. 61 (0) 2 8877 9101
www.btfbio.com



Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
 69280 Marcy l'Etoile - France
 Tel. 33 (0) 4 78 87 20 00
 Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

