

**Batch Number:** **B 666**

**Reference Material Data**

Organism: *Salmonella abony*  
Designation: ACM 5080 (NCTC 6017)  
Catalogue No: B-SA5080-550-20  
Date of Manufacture: 6 July 2006  
Expiry Date: 5 July 2008  
Extended Expiry Date<sup>#</sup>: 28 April 2010

**Quantification Data**

(Method Ref: BB-019)

**cfu Count**

Mean cfu	SD
<b>517.7</b>	<b>26.5</b>

**Culture Information**

No. of Passages: 4  
QC Media: Nutrient Agar  
Plate Incubation Time (hrs): 14 - 24  
Plate Incubation Temp °C: 37

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing.

Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Senior Scientist:**

Charlotte Morgan

Strain sourced from ACM (Australian Collection of Microorganisms)

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall<sup>TM</sup> products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall<sup>TM</sup> batch has been extended.

