

**Batch Number: B 809**

**Reference Material Data**

Organism: *Streptococcus pyogenes*  
Designation: NCTC 12696/ATCC 19615  
Catalogue No: B-SP12696-550  
Date of Manufacture: 15 March 2007  
Expiry Date: 14 March 2009  
Extended Expiry Date#: N/A

**Quantification Data**

(Method Ref: BB-019)

**cfu Count**

Mean cfu	SD
<b>532.7</b>	<b>50.2</b>

**Culture Information**

No. of Passages: 4  
QC Media: Nutrient agar  
Plate Incubation Time (hrs): 24 - 36  
Plate Incubation Temp °C: 37

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed using the Dupont Qualicon Riboprinting method of genetic typing.

Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Chief Technical Officer:**



.....  
Graham Vesey



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC  
ATCC is a trademark of American Type Culture Collection

\*BTF regularly extends the expiry dates on standard BioBall™ products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall™ batch has been extended.

