

Batch Number: **B 846**

Reference Material Data

Organism: *Escherichia coli*
Designation: NCTC 12241 (ATCC 25922)
Catalogue No: B-EC12241-5-20
Date of Manufacture: 19 June 2007
Expiry Date: 18 June 2009
Extended Expiry Date[#]: 8 December 2010

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Chief Technical Officer:



.....
Graham Vesey

Water Industry Enquires:

BTF Pty Ltd.
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com

Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

Quantification Data

(Method Ref: BB-001)

cfu Count

Mean cfu	SD
5.6	1.0

95% Prediction Interval*

Lower Limit	Upper Limit
3.5	7.7

*For an explanation of the Prediction Interval see the BioBall® Instructions for Use on the BTF website.



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)
ATCC is a trademark of American Type Culture Collection

[#]BTF regularly extends the expiry dates on standard BioBall® products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall® batch has been extended.



A BIOMÉRIEUX COMPANY

