

Batch Number: B 906

Reference Material Data

Organism: *Pseudomonas aeruginosa*
Designation: NCTC 12924/ATCC 9027
Catalogue No: B-PA12924-30
Date of Manufacture: 4 October 2007
Expiry Date: 3 October 2009
Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-001)

cfu Count

Mean cfu	SD
31.1	2.8

95% Prediction Interval*

Lower Limit	Upper Limit
25.4	36.8

*See the BioBall™ quality control information on the product insert for an explanation of the prediction interval

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed using the Dupont Qualicon Ribotyping method of genetic typing.

Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Chief Technical Officer:



.....
Graham Vesey



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC

ATCC is a trademark of American Type Culture Collection

*BTF regularly extends the expiry dates on standard BioBall™ products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall™ batch has been extended.



BTF Pty Ltd | ACN 086488084 | ABN 22086488084 | PO Box 599 North Ryde BC NSW 1670 Sydney Australia
Phone: +61 2 8877 9150 | Fax: +61 2 8877 9101 | Email: info@btfbio.com | www.btfbio.com