

Batch Number: **B 909****Reference Material Data**

Organism: *Escherichia coli*  
Designation: NCTC 12241 (ATCC 25922)  
Catalogue No: B-EC12241-30-20  
Date of Manufacture: 10 October 2007  
Expiry Date: 9 October 2009  
Extended Expiry Date<sup>#</sup>: 8 March 2011

**Culture Information**

No. of Passages: 4  
QC Media: Nutrient agar  
Plate Incubation Time (hrs): 14 - 24  
Plate Incubation Temp °C: 37

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Chief Technical Officer:**

.....  
Graham Vesey

**Water Industry Enquires:**

BTF Pty Ltd.  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101

[www.btfbio.com](http://www.btfbio.com)

**Food & Pharmaceutical Industry Enquires:**

bioMérieux S.A  
69280 Marcy l'Etoile - France  
Tel. 33 (0) 4 78 87 20 00  
Fax. 33 (0) 4 78 87 20 90

[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

**Quantification Data**

(Method Ref: BB-001)

**cfu Count**

| Mean cfu    | SD         |
|-------------|------------|
| <b>30.6</b> | <b>3.3</b> |

**95% Prediction Interval\***

| Lower Limit | Upper Limit |
|-------------|-------------|
| <b>23.8</b> | <b>37.3</b> |

\*For an explanation of the Prediction Interval see the BioBall<sup>®</sup> Instructions for Use on the BTF website.



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)  
ATCC is a trademark of American Type Culture Collection

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall<sup>®</sup> products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall<sup>®</sup> batch has been extended.



A BIOMÉRIEUX COMPANY

