

Batch Number: B 933

Reference Material Data

Organism: *Enterococcus faecalis*
Designation: NCTC 775/ATCC 19433
Catalogue No: B-EF775-30
Date of Manufacture: 21 November 2007
Expiry Date: 20 November 2009
Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-001)

cfu Count

Mean cfu	SD
30.1	2.5

95% Prediction Interval*

Lower Limit	Upper Limit
25.0	35.2

*See the BioBall™ quality control information on the product insert for an explanation of the prediction interval

Culture Information

No. of Passages: 4
QC Media: Horse Blood agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed using the Dupont Qualicon Ribotyping method of genetic typing.

Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Chief Technical Officer:



.....
Graham Vesey



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC

ATCC is a trademark of American Type Culture Collection

*BTF regularly extends the expiry dates on standard BioBall™ products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall™ batch has been extended.



PRECISE
MICROBIOLOGY

BTF Pty Ltd | ACN 086488084 | ABN 22086488084 | PO Box 599 North Ryde BC NSW 1670 Sydney Australia
Phone: +61 2 8877 9150 | Fax: +61 2 8877 9101 | Email: info@btfbio.com | www.btfbio.com