

**Batch Number: B 941**

### Reference Material Data

Organism: *Staphylococcus aureus*  
Designation: NCTC 10788/ATCC 6538  
Catalogue No: B-SA10788-30  
Date of Manufacture: 4 December 2007  
Expiry Date: 3 December 2009  
Extended Expiry Date#: N/A

### Quantification Data

(Method Ref: BB-001)

#### cfu Count

Mean cfu	SD
31.0	2.0

#### 95% Prediction Interval\*

Lower Limit	Upper Limit
26.9	35.1

\*See the BioBall™ quality control information on the product insert for an explanation of the prediction interval

### Culture Information

No. of Passages: 4  
QC Media: Nutrient agar  
Plate Incubation Time (hrs): 14 - 24  
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

### Strain Confirmation

Bacteria strains are confirmed using the Dupont Qualicon Ribotyping method of genetic typing.

Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

### Chief Technical Officer:



.....  
Graham Vesey



This document is issued in accordance with NATA's accreditation requirements. Accredited for compliance with ISO Guide 34.

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC

ATCC is a trademark of American Type Culture Collection

\*BTF regularly extends the expiry dates on standard BioBall™ products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall™ batch has been extended.



BTF Pty Ltd | ACN 086488084 | ABN 22086488084 | PO Box 599 North Ryde BC NSW 1670 Sydney Australia  
Phone: +61 2 8877 9150 | Fax: +61 2 8877 9101 | Email: [info@btfbio.com](mailto:info@btfbio.com) | [www.btfbio.com](http://www.btfbio.com)