

Batch Number: B 804

Reference Material Data

Organism: *Listeria monocytogenes*
Designation: BTF 161
Catalogue No: B-GLM161-30
Date of Manufacture: 13 March 2007
Expiry Date: 12 March 2009
Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-001)

cfu Count

Mean cfu	SD
28.1	3.0

95% Prediction Interval*

Lower Limit	Upper Limit
22.2	34.0

*See the BioBall™ quality control information on the product insert for an explanation of the prediction interval

Culture Information

No. of Passages: 4
QC Media: Horse Blood agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Senior Scientist:



Charlotte Morgan

#BTF regularly extends the expiry dates on standard BioBall™ products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall™ batch has been extended.

Regulations and laws controlling Genetically Modified Organisms (GMOs)
The fluorescent bacteria in BioBall™ Glo are genetically modified to express a modified green fluorescent protein derived from a jellyfish *Aequorea Victoria*.

It is the responsibility of the user to ensure that regulations and laws regarding the handling of GMOs are followed and that all necessary permits are obtained.